



Group Report
Round 8A

Issue 2



November 2018

Scanning Electron Microscopy Scheme

This is a re-issued report and supersedes the SEMS Group Report issued on 9th November 2018.
Reason for Re-Issue: Error in data supplied in Appendix 1.

BACKGROUND

This report covers Round 8A of the SEMS asbestos fibre counting PT scheme. The scheme is operated by HSL, in collaboration with APC, Germany and TNO, Netherlands.

SAMPLES

Four samples were circulated representing a range of different fibre densities and fibre types. All samples were produced at HSL using the modified sputnik multi-port sampling instrument.

INTRODUCTION

A total of 53 laboratories participated in this round (including the validating laboratories). Laboratories were able to submit up to three results per sample and many laboratories took advantage of this with a total of 390 results submitted.

The samples were as follows:

8ASEM1 - Low density (0 fibres/mm²) - no asbestos fibres

8ASEM2 - Medium density (17.1 fibres/mm²) - amosite fibres

8ASEM3 - Medium density (29.0 fibres/mm²) - amosite & crocidolite fibres

8ASEM4 - Medium density (16.4 fibres/mm²) - chrysotile fibres

INFORMATION SUBMITTED BY LABORATORIES

Laboratories were asked to supply the following information:



Group Report
Round 8A

Issue 2



November 2018

Scanning Electron Microscopy Scheme

- Number of fibres $>5\mu\text{m}$ in length counted (amphibole, chrysotile & other inorganic)
- The number of fields of view searched
- The area of the field of view
- The magnification and the method used

Laboratories were asked to calculate the fibre density (in fibres/ mm^2) for each fibre type identified. There was also an option to include the number of fibres $\leq 5\mu\text{m}$ in length.

LABORATORY ASSESSMENT

RESULTS

Calculations – No errors were identified in this round.

Screen area – The fibre densities submitted by laboratories have not been recalculated and the density calculation and therefore screen area has not been verified.

Magnification – As was the case in earlier rounds, some laboratories used an operating magnification outside the range defined in ISO 14966 (or VDI 3492).

Magnifications of 4000x, 3000x, 2000x and 1000x were recorded.

Results for total asbestos fibre densities for each laboratory are summarised in Appendix 1.

Data Analysis

Data analysis is based upon the total asbestos fibre densities (amphibole & chrysotile) derived from fibre numbers counted and the area of the filter searched. The distribution of fibres on a filter derived from airborne sampling is normally described as being Poisson-distributed. For Poisson-distributed counts, the variance (standard deviation squared) is equal to the mean. However, in practice the variation may be larger due to differences in sample production, laboratories and individual microscopists.

A comparison of the observed standard deviations with the expected standard deviations (expected under Poisson distribution) show that the observed variation is larger than that expected, and it is difficult to quantify how much of this may be due to differences in sample production, and how much is due to differences between labs/microscopists.

For this report, the data have been compared against the criteria used in the UK phase contrast fibre counting proficiency testing scheme RICE. Details of the analysis used can be found in Appendix 2.

Round 8A Overview

Summary statistics from this round of results are displayed in Table 1. Below this, Figure 1 displays the percentage of participants in each scoring band (as per the RICE scoring system). Figures 2 and 3 show the band scored by participants divided according to magnification and method used respectively.

Table 1: Summary statistics for results received in SEMS Round 8A.

| | Sample 1 | Sample 2 | Sample 3 | Sample 4 |
|--|----------|----------|----------|----------|
| Number of results | 96 | 97 | 99 | 98 |
| Median (fibres/mm²) | 0.0 | 17.1 | 29.0 | 16.4 |
| 25th percentile (fibres/mm²) | 0.0 | 14.1 | 22.3 | 11.1 |
| 75th percentile (fibres/mm²) | 0.0 | 23.0 | 39.6 | 25.9 |
| Interquartile range (fibres/mm²) | 0.0 | 8.9 | 17.3 | 14.8 |
| Mean (fibres/mm²) | 1.8 | 18.9 | 33.9 | 21.1 |
| Standard deviation (fibres/mm²) | 8.7 | 9.0 | 19.0 | 16.8 |
| Relative standard deviation (%) | 498.7 | 47.5 | 56.1 | 79.7 |

*Note: The relative standard deviation (RSD) is calculated by (standard deviation/mean)*100%. This statistic illustrates the variation relative to the size of the mean value. For very low values of the mean (e.g. Sample 1), the value of the RSD can be considered largely meaningless.*




Figure 1: Banded scores for participants in SEMS Round 8A (categorised as per RICE scoring system - see Appendix 2)




Figure 2: Banded scores for participants in SEMS Round 8A divided according to method used




Figure 3: Banded scores for participants in SEMS Round 8A divided according to magnification used

APPENDIX 1

Sample 1 (8ASEM1) - Low density (0 fibres mm⁻²) - no asbestos fibres

| LAB NUMBER | TOTAL ASBESTOS | BAND (RICE) |
|------------|----------------|-------------|
| 1277 | 0 | A |
| 1277 | 0 | A |
| 1620 | 0 | A |
| 1620 | 0 | A |
| 1620 | 0 | A |
| 1831 | 0 | A |
| 1831 | 0 | A |
| 1831 | 0 | A |
| 1832 | 0 | A |
| 1836 | 66.59 | C |
| 1848 | 0 | A |
| 1852 | 0 | A |
| 1860 | 6.61 | B |
| 1866 | 0 | A |
| 1866 | 0 | A |
| 1868 | 0 | A |
| 1868 | 0 | A |
| 1875 | 0 | A |
| 1877 | 0 | A |
| 1879 | 0 | A |
| 1879 | 0 | A |
| 1879 | 0 | A |
| 1889 | 0 | A |
| 1889 | 0 | A |
| 1903 | 0 | A |
| 1910 | 0 | A |
| 1922 | 0 | A |
| 1922 | 0 | A |
| 1922 | 0 | A |

APPENDIX 1

| | | |
|------|-------|---|
| 1923 | 0 | A |
| 1923 | 0 | A |
| 1928 | 0 | A |
| 1928 | 0 | A |
| 1928 | 0 | A |
| 1936 | 0 | A |
| 1936 | 0 | A |
| 1937 | 0 | A |
| 1937 | 0 | A |
| 1938 | 0 | A |
| 1939 | 0 | A |
| 1939 | 0 | A |
| 1940 | 0 | A |
| 1948 | 0 | A |
| 1958 | 0 | A |
| 1966 | 0 | A |
| 1966 | 0 | A |
| 1966 | 0 | A |
| 1968 | 0 | A |
| 1976 | 0 | A |
| 1976 | 0 | A |
| 1977 | 0 | A |
| 1981 | 0 | A |
| 1981 | 0 | A |
| 1992 | 0 | A |
| 1992 | 0 | A |
| 1993 | 0 | A |
| 1993 | 0 | A |
| 1993 | 0 | A |
| 1999 | 26.12 | C |
| 1999 | 31.54 | C |
| 1999 | 37.45 | C |

APPENDIX 1

| | | |
|------|---|---|
| 2020 | 0 | A |
| 2020 | 0 | A |
| 2020 | 0 | A |
| 2022 | 0 | A |
| 2023 | 0 | A |
| 2023 | 0 | A |
| 2023 | 0 | A |
| 2024 | 0 | A |
| 2024 | 0 | A |
| 2026 | 0 | A |
| 2026 | 0 | A |
| 2026 | 0 | A |
| 2037 | 0 | A |
| 2044 | 0 | A |
| 2044 | 0 | A |
| 2044 | 0 | A |
| 2051 | 0 | A |
| 2061 | 0 | A |
| 2061 | 0 | A |
| 2061 | 0 | A |
| 2063 | 0 | A |
| 2069 | 0 | A |
| 2076 | 0 | A |
| 2076 | 0 | A |
| 2076 | 0 | A |
| 2098 | 0 | A |
| 2107 | 0 | A |
| 2107 | 0 | A |
| 2108 | 0 | A |
| 2110 | 0 | A |
| 2116 | 0 | A |
| 2132 | 0 | A |

APPENDIX 1

| | | |
|------|---|---|
| 2132 | 0 | A |
| 2132 | 0 | A |
| 2135 | 0 | A |

| | |
|---------------------|------------|
| Mean | 0.0 |
| Median (Ref) | 0.0 |
| STDev | 8.7 |
| Min | 0.0 |
| Max | 66.59 |

| | | | | | |
|---------|---------|---------|---------|---------|---------|
| RICE A | RICE A | RICE B | RICE B | RICE C | RICE C |
| (Lower) | (Upper) | (Lower) | (Upper) | (Lower) | (Upper) |
| - | 3.8 | - | 10.9 | - | >10.9 |

Sample 2 (8ASEM2) - Medium density (17.1 fibres/mm²) - amosite fibres

| LAB NUMBER | TOTAL ASBESTOS | BAND (RICE) |
|------------|----------------|-------------|
| 1277 | 15.8 | A |
| 1277 | 17.5 | A |
| 1620 | 16.0 | A |
| 1620 | 17.5 | A |
| 1620 | 20.0 | A |
| 1831 | 14.1 | A |
| 1831 | 14.4 | A |
| 1831 | 14.8 | A |
| 1832 | 11.0 | A |
| 1836 | 52.7 | B |
| 1848 | 12.8 | A |
| 1852 | 17.3 | A |
| 1860 | 9.4 | A |
| 1866 | 12.6 | A |
| 1866 | 20.6 | A |
| 1868 | 15.5 | A |
| 1868 | 16.4 | A |
| 1875 | 18.3 | A |
| 1877 | 24.0 | A |
| 1879 | 7.0 | A |
| 1879 | 10.0 | A |
| 1879 | 12.0 | A |
| 1889 | 17.1 | A |
| 1889 | 29.1 | A |
| 1903 | 20.9 | A |
| 1903 | 25.9 | A |
| 1910 | 16.0 | A |
| 1922 | 16.7 | A |
| 1922 | 23.6 | A |

APPENDIX 1

| | | |
|------|------|---|
| 1922 | 30.9 | A |
| 1923 | 21.8 | A |
| 1923 | 24.5 | A |
| 1928 | 10.7 | A |
| 1928 | 12.7 | A |
| 1928 | 14.1 | A |
| 1936 | 31.0 | A |
| 1936 | 32.0 | A |
| 1937 | 13.3 | A |
| 1937 | 14.8 | A |
| 1938 | 51.0 | B |
| 1939 | 17.5 | A |
| 1939 | 21.0 | A |
| 1940 | 24.2 | A |
| 1948 | 15.0 | A |
| 1958 | 17.8 | A |
| 1966 | 15.8 | A |
| 1966 | 16.8 | A |
| 1966 | 16.8 | A |
| 1968 | 14.0 | A |
| 1976 | 15.0 | A |
| 1976 | 18.0 | A |
| 1977 | 19.7 | A |
| 1981 | 1.9 | C |
| 1981 | 6.7 | A |
| 1992 | 34.1 | A |
| 1992 | 39.2 | B |
| 1993 | 4.0 | B |
| 1993 | 6.0 | B |
| 1993 | 20.0 | A |
| 1999 | 25.1 | A |
| 1999 | 26.6 | A |

APPENDIX 1

| | | |
|------|------|---|
| 1999 | 33.5 | A |
| 2020 | 20.9 | A |
| 2020 | 21.1 | A |
| 2020 | 21.2 | A |
| 2022 | 15.0 | A |
| 2023 | 12.3 | A |
| 2023 | 17.1 | A |
| 2023 | 22.8 | A |
| 2024 | 0.2 | C |
| 2024 | 0.2 | C |
| 2026 | 21.2 | A |
| 2026 | 23.7 | A |
| 2026 | 30.6 | A |
| 2037 | 16.7 | A |
| 2044 | 19.4 | A |
| 2044 | 26.6 | A |
| 2044 | 35.5 | A |
| 2051 | 16.1 | A |
| 2061 | 11.4 | A |
| 2061 | 12.3 | A |
| 2061 | 15.3 | A |
| 2063 | 23.0 | A |
| 2069 | 33.0 | A |
| 2076 | 19.0 | A |
| 2076 | 25.0 | A |
| 2076 | 33.0 | A |
| 2098 | 17.0 | A |
| 2107 | 13.0 | A |
| 2107 | 15.0 | A |
| 2108 | 16.5 | A |
| 2110 | 9.8 | A |
| 2116 | 19.0 | A |

APPENDIX 1

| | | |
|------|------|---|
| 2132 | 12.9 | A |
| 2132 | 13.2 | A |
| 2132 | 15.0 | A |
| 2135 | 29.0 | A |

Mean 18.9

Median 17.1

STDev 9.0

Min 0.2

Max 52.7

| | | | | | |
|----------------|----------------|----------------|----------------|----------------|----------------|
| RICE A | RICE A | RICE B | RICE B | RICE C | RICE C |
| (Lower) | (Upper) | (Lower) | (Upper) | (Lower) | (Upper) |
| 6.6 | 37.2 | 3.2 | 55.3 | <6.6 | >55.3 |

Sample 3 (8ASEM3) - Medium density (29.0 fibres/mm²) - amosite & crocidolite fibres fibres

| LAB NUMBER | TOTAL ASBESTOS | BAND (RICE) |
|------------|----------------|-------------|
| 1277 | 23.2 | A |
| 1277 | 26.7 | A |
| 1620 | 26.0 | A |
| 1620 | 27.0 | A |
| 1620 | 34.5 | A |
| 1831 | 16.1 | A |
| 1831 | 17.1 | A |
| 1831 | 17.8 | A |
| 1832 | 22.0 | A |
| 1836 | 87.5 | C |
| 1848 | 28.3 | A |
| 1852 | 38.3 | A |
| 1860 | 13.2 | B |
| 1866 | 15.4 | A |
| 1866 | 18.2 | A |
| 1868 | 37.3 | A |
| 1868 | 39.1 | A |
| 1875 | 36.6 | A |
| 1877 | 46.0 | A |
| 1879 | 26.0 | A |
| 1879 | 29.0 | A |
| 1879 | 29.0 | A |
| 1889 | 40.2 | A |
| 1889 | 44.2 | A |
| 1903 | 38.9 | A |
| 1903 | 46.8 | A |
| 1903 | 57.8 | B |
| 1910 | 63.0 | B |
| 1922 | 17.7 | A |

APPENDIX 1

| | | |
|------|------|---|
| 1922 | 20.6 | A |
| 1922 | 32.4 | A |
| 1923 | 26.1 | A |
| 1923 | 32.5 | A |
| 1928 | 20.1 | A |
| 1928 | 20.8 | A |
| 1928 | 20.8 | A |
| 1936 | 9.0 | C |
| 1936 | 10.0 | B |
| 1937 | 18.1 | A |
| 1937 | 21.4 | A |
| 1938 | 35.0 | A |
| 1938 | 37.0 | A |
| 1939 | 23.0 | A |
| 1939 | 26.5 | A |
| 1940 | 37.5 | A |
| 1948 | 22.5 | A |
| 1958 | 21.9 | A |
| 1966 | 23.7 | A |
| 1966 | 23.7 | A |
| 1966 | 27.3 | A |
| 1968 | 20.0 | A |
| 1976 | 28.0 | A |
| 1976 | 28.0 | A |
| 1977 | 26.5 | A |
| 1981 | 10.5 | B |
| 1981 | 12.4 | B |
| 1992 | 33.1 | A |
| 1992 | 62.2 | B |
| 1993 | 43.0 | A |
| 1993 | 68.0 | B |
| 1993 | 83.0 | C |

APPENDIX 1

| | | |
|------|------|---|
| 1999 | 81.8 | C |
| 1999 | 86.2 | C |
| 1999 | 96.1 | C |
| 2020 | 29.5 | A |
| 2020 | 31.4 | A |
| 2020 | 32.7 | A |
| 2022 | 24.0 | A |
| 2023 | 30.4 | A |
| 2023 | 37.9 | A |
| 2023 | 38.9 | A |
| 2024 | 0.2 | C |
| 2024 | 0.3 | C |
| 2026 | 48.3 | A |
| 2026 | 53.2 | A |
| 2026 | 68.0 | B |
| 2037 | 23.3 | A |
| 2044 | 24.7 | A |
| 2044 | 27.6 | A |
| 2044 | 88.3 | C |
| 2051 | 25.0 | A |
| 2061 | 21.7 | A |
| 2061 | 31.1 | A |
| 2061 | 31.1 | A |
| 2063 | 40.0 | A |
| 2069 | 51.0 | A |
| 2076 | 44.0 | A |
| 2076 | 48.0 | A |
| 2076 | 49.0 | A |
| 2098 | 49.5 | A |
| 2107 | 24.0 | A |
| 2107 | 38.0 | A |
| 2108 | 11.5 | B |

APPENDIX 1

| | | |
|------|------|---|
| 2110 | 10.8 | B |
| 2116 | 38.0 | A |
| 2132 | 24.3 | A |
| 2132 | 30.4 | A |
| 2132 | 36.1 | A |
| 2135 | 41.0 | A |

| | |
|---------------------|------|
| Mean | 33.9 |
| Median (Ref) | 29.0 |
| STDev | 19.0 |
| Min | 0.2 |
| Max | 96.1 |

| | | | | | |
|----------------|----------------|----------------|----------------|----------------|----------------|
| RICE A | RICE A | RICE B | RICE B | RICE C | RICE C |
| (Lower) | (Upper) | (Lower) | (Upper) | (Lower) | (Upper) |
| 14.6 | 54.0 | 9.3 | 75.4 | <9.3 | >75.4 |

Sample 4 (8ASEM4) - Medium density (16.4 fibres/mm²) - chrysotile fibres

| LAB NUMBER | TOTAL ASBESTOS | BAND (RICE) |
|------------|----------------|-------------|
| 1277 | 5.8 | B |
| 1277 | 10.8 | A |
| 1620 | 13.0 | A |
| 1620 | 16.5 | A |
| 1620 | 18.0 | A |
| 1831 | 11.4 | A |
| 1831 | 14.1 | A |
| 1831 | 15.8 | A |
| 1832 | 6.0 | B |
| 1836 | 64.6 | C |
| 1848 | 16.0 | A |
| 1852 | 11.9 | A |
| 1860 | 5.7 | B |
| 1866 | 22.4 | A |
| 1866 | 25.2 | A |
| 1868 | 14.5 | A |
| 1868 | 18.2 | A |
| 1875 | 10.8 | A |
| 1877 | 38.0 | B |
| 1879 | 13.0 | A |
| 1879 | 13.0 | A |
| 1879 | 14.0 | A |
| 1889 | 21.1 | A |
| 1889 | 22.1 | A |
| 1903 | 41.8 | B |
| 1903 | 51.8 | B |
| 1910 | 9.0 | A |
| 1922 | 20.6 | A |
| 1922 | 27.0 | A |

APPENDIX 1

| | | |
|------|-------|---|
| 1922 | 29.9 | A |
| 1923 | 4.8 | B |
| 1923 | 5.9 | B |
| 1928 | 14.7 | A |
| 1928 | 15.4 | A |
| 1928 | 18.1 | A |
| 1936 | 7.0 | A |
| 1936 | 11.0 | A |
| 1937 | 6.7 | A |
| 1937 | 6.7 | A |
| 1938 | 12.0 | A |
| 1938 | 30.0 | A |
| 1939 | 10.0 | A |
| 1939 | 14.0 | A |
| 1940 | 30.0 | A |
| 1948 | 11.0 | A |
| 1958 | 9.1 | A |
| 1966 | 11.9 | A |
| 1966 | 12.8 | A |
| 1966 | 14.8 | A |
| 1968 | 15.0 | A |
| 1976 | 22.0 | A |
| 1976 | 25.0 | A |
| 1977 | 19.0 | A |
| 1981 | 0.0 | C |
| 1981 | 1.9 | C |
| 1992 | 3.0 | B |
| 1992 | 3.0 | B |
| 1993 | 26.0 | A |
| 1993 | 35.0 | A |
| 1993 | 106.0 | C |
| 1999 | 56.7 | C |

APPENDIX 1

| | | |
|------|------|---|
| 1999 | 59.6 | C |
| 1999 | 68.5 | C |
| 2020 | 19.2 | A |
| 2020 | 21.6 | A |
| 2020 | 22.4 | A |
| 2022 | 14.0 | A |
| 2023 | 16.2 | A |
| 2023 | 22.8 | A |
| 2023 | 27.5 | A |
| 2024 | 0.1 | C |
| 2024 | 0.2 | C |
| 2026 | 16.3 | A |
| 2026 | 25.6 | A |
| 2026 | 26.6 | A |
| 2037 | 16.7 | A |
| 2044 | 9.7 | A |
| 2044 | 14.5 | A |
| 2044 | 15.5 | A |
| 2051 | 17.9 | A |
| 2061 | 36.6 | B |
| 2061 | 46.9 | B |
| 2061 | 48.9 | B |
| 2063 | 22.0 | A |
| 2069 | 18.0 | A |
| 2076 | 36.0 | A |
| 2076 | 44.0 | B |
| 2076 | 47.0 | B |
| 2098 | 25.0 | A |
| 2107 | 11.0 | A |
| 2107 | 14.0 | A |
| 2108 | 0.0 | C |
| 2110 | 0.0 | C |

APPENDIX 1

| | | |
|------|------|---|
| 2116 | 21.0 | A |
| 2132 | 24.6 | A |
| 2132 | 26.2 | A |
| 2132 | 35.0 | A |
| 2135 | 40.0 | B |

| | |
|---------------------|-------|
| Mean | 21.1 |
| Median (Ref) | 16.4 |
| STDev | 16.8 |
| Min | 0.0 |
| Max | 106.0 |

| | | | | | |
|----------------|----------------|----------------|----------------|----------------|----------------|
| RICE A | RICE A | RICE B | RICE B | RICE C | RICE C |
| (Lower) | (Upper) | (Lower) | (Upper) | (Lower) | (Upper) |
| 6.1 | 36.1 | 2.9 | 54.0 | <6.1 | >54.0 |

APPENDIX 2

DATA ANALYSIS

Regular Inter-laboratory Counting Exchange (RICE) Criteria

Where R is the reference value – in this case the Median value.

High density samples ($R > 63.7$ fibres. mm^{-2})

Target band A: $> 0.65R$ to $< 1.55R$

Target band B: $> 0.50R$ to $0.65R$ [band -B] and $> 1.55R$ to $2.00R$ [band +B]

Target band C: $< 0.50R$ [band -C] and $> 2.00R$ [band +C]

Low density samples ($R \leq 63.7$ fibres. mm^{-2})*

Target band A: $(\sqrt{R-1.57})^2$ to $(\sqrt{R+1.96})^2$ [band A]

Target band B: $< (\sqrt{R-2.34})^2$ to $(\sqrt{R-1.57})^2$ [band -B]
 $> (\sqrt{R+1.96})^2$ to $(\sqrt{R+3.30})^2$ [band +B]

Target band C: $< (\sqrt{R-2.34})^2$ [band -C]
 $> (\sqrt{R+3.30})^2$ [band +C]

* For samples less than 5.5 fibres. mm^{-2} the lower limit is set to zero when the component within the brackets $(\sqrt{R-n})$ is less than zero.

The plot below shows the positions of the performance limits in relation to the reference counts up to reference density 500 fibres per mm^2 .

